

GH+
Labs

Sample Prep Module | NAAATOS
kickoff
—

Feb 14, 2024



Agenda

Intros	All
Overview / how we got here	Matt K
Deeper look at specs	Matt K
Current design details	Bryan N
Plan moving forward	All




Overview, or how did we get here



Many alternative TB testing methods, tongue swab is the best

#	Category	Specimen	Benefits of testing method	Limitations
1	Current Method	Sputum	<ul style="list-style-type: none"> Used for 140 years to diagnose TB Can usually be obtained from patients with respiratory symptoms, given enough time/effort Usable for diagnosis, treatment monitoring, and DST Current DST ecosystem designed around sputum 	<ul style="list-style-type: none"> Unwelcome & stigmatized Difficult to obtain in children, PLHIV, elderly, and in systematic screening Occupational hazard for HCW Viscous, heterogenous, poorly reproducible, difficult to process
2	Alternative Respiratory-based testing methods	Tongue Swab	<ul style="list-style-type: none"> The world just got swabbed; scalable TB is recoverable from tongue swabs with good concordance with sputum results Tongue swab eluent offers highly-simplified specimen processing, relative to sputum Highly stable specimen, even at elevated temps 	<ul style="list-style-type: none"> Less sensitive for very low bacillary loads
3		Cough		
4		Saliva	<ul style="list-style-type: none"> Easy to test 	<ul style="list-style-type: none"> Larger volume required Dilute sample
5		Breath	<ul style="list-style-type: none"> Easy to test 	<ul style="list-style-type: none"> Early stages of product development, still unclear what results mean
6	Alternative Non-respiratory based methods	LAM	<ul style="list-style-type: none"> Simplified sampling, compared to sputum May have extrapulmonary TB Dx advantages Low cost 	<ul style="list-style-type: none"> Suboptimal diagnostic accuracy (LAM)
7		Stool (peds) Host RNA Cell-free NA	<ul style="list-style-type: none"> Simplified sampling, compared to sputum Enables sampling of pediatric populations May be able to diagnose subclinical disease 	<ul style="list-style-type: none"> Complicated sample handling (stool) Early stages of specimen characterization (RNA and cfDNA) Complicated processing (RNA, cell-free)

Current tongue swab consortium “standard” sample prep accomplishes 3 primary tasks we need to translate for POC



The fortified TB envelope complicates approaches to lysis





Image from: Van Der Horst. *Analyst*. 2021.




Current lysis approach characteristics and limitations informed the design space for novel mechanical BB solutions at POC



	Biospec Mini-Beadbeater	Vortexer + Adapter	SoniBeast Jr	Lego hand-crank	Hand-shaking
Energy Requirement	Electrical Power			No Electrical Required	
Oscillation frequency (Hz)	58	50	300	10-15	2-3
Throw / displacement (mm)	19	4+	2.5	25	>100
Motion profile	Figure 8	Circle	Linear oscillation	Elliptical	Hammer-like
Power source	Mains	Mains	Lithium battery	Manual	Manual
Power draw (W)	850	135	50	N/A	N/A
Time for optimal result (min)	3 x 1	5	2	Not possible	>5*
User experience	Walkaway	Walkaway	Uncomfortable	Variable / tiring	Variable / tiring
Price (\$)	2500	600	320	10	0



Why we focused on voice coils for internal development

- Considered many ways to shake a tube within bounds of previous slide, and settled on voice coils as a cheap, rugged technology both to study key motion parameters and potentially become a product
- Relative to many other mechanical means, voice coils offer:
 - High efficiency due to their harmonic nature, which allows linear motion without the need for constant acceleration and deceleration of the load
 - No contact wear or friction given their structure, which also minimizes unwanted vibrations / sounds
 - Extreme longevity without maintenance given the above and the long history of commodity-level production
- Primary challenge to overcome for use of voice coils was adapting the flexure plate to support the relatively larger displacements needed for lysis compared with, e.g. vibrating a speaker cone





Several key learnings stemmed from initial voice coil work

- General experimental notes:
 - All results noted here used TB-spiked oral matrix (from volunteer tongue swabs) and 150mg of 100 μ m glass beads in the sample tube
 - Lysis efficiency (%) calculated relative to the copies/mL PCR output from the TB tongue swab reference assay of 10 minutes at 95C and bead beating in the BioSpec mini bead beater
- Lysis was dramatically more efficient when shaking perpendicular to the tube's long axis compared to parallel
- Orienting the tube horizontally vs. vertically (upright) did not have any notable effects
- Leaving the swab tip in the tube equaled or improved lysis efficiency vs. no swab tip
- Bead type, size, and amount maintained their parameter values from previous optimization for use in other mechanical bead beating systems
- Initially performed the bead beating after letting samples cool to room temp following 10 minutes at 95C
 - In this condition, optimal lysis efficiency required 8-9 mm displacement, or throw
 - Frequency had a smaller effect than throw, with sufficient performance in the 70-110 Hz range



Lysis becomes easier with integrated heater


- Initial experiment showed ~10% lysis efficiency improvement when putting samples directly into voice coil setup from the heat block
- Devised new architecture to integrate the heater into the voice coil tube holder
 - Both the heater and tube holder are constructed from a stack of printed circuit boards (PCBs)
 - Double duty keeps the heater/tube holder cheap and light
 - Found we can heat and shake 4 tubes at once using a single voice coil / flexure plate
- New integrated design showed substantially higher lysis efficiency compared with bead beating room temp samples (fig on right)
 - No significant differences between leaving heater on during shaking and turning off heater just before shaking
 - PCB heater/tube holder insulates tubes well against heat loss
- Integrated heater allowed for smaller displacements (~7mm) and lower frequencies (50-70Hz) to achieve optimal lysis efficiency
- Subsequent optimizations of temperature, heating duration, and bead beating duration for lysis efficiency, nuclease inhibition, and biosafety led to the “final” protocol on following slide




Current Design Approach | Reusable Lysis Module

- Current plan to meet product requirements is our voice coil bead beater (VCBB) with integrated heating
 - Lysis performance at least on par with “gold standard” heat plus BioSpec bead beating (detailed on next slide)
 - Run time of 6-7 minutes for 1-4 tubes at a time
 - Temp ramps to 95C in ~1-2 minutes, depending on starting condition
 - Temp holds at 95C for 2 minutes, then heater turns off
 - Voice coil bead beating lasts for 4 minutes, sample remains hot during that time to aid lysis performance
 - Projected BOM < \$100 at moderate scale of quantity 1000
 - Minimal power draw to enable easy battery operation
- Prototypes have gone through several iterations, will soon undergo external DfM and productization


First heat-integrated lab version



Uganda study (R1)



Further lab studies (R2)




Final GH+Labs rev (R3)




VCBB has demonstrated performance in line with reference assays

- Uganda R2D2 site study with patient samples confirmed previous findings for VCBB R1 with contrived samples
 - VCBB R1 typically outperformed BioSpec BB, except at very low copy numbers (Rev3 design alleviates this issue)
- Lysis efficiency is maintained for temps down to 65C, but nuclease inactivation is not as effective at lower temp
 - Tested by letting samples sit at room temp for 60 mins following integrated heat and lysis, before PCR



- Using the Uganda study unit and protocol, biosafety experiments showed 0 colonies at 8 weeks




No treatment

VCBB

Continuing to support and monitor other sonication sample prep approaches

- Performed explorations with an ultrasonic horn OEM dev kit
 - Very similar to what is used in GeneXpert
 - Explored lysis efficiency as function of:
 - Ultrasonic horn amplitude
 - Clamping force
 - Tube material
 - Duration (total and per cycle)
 - Presence and amount of beads
- Discovered that conditions we found to be optimal were very similar to a collaborating company
 - Two cycles of 30s with 30s off between, include glass beads, draw ~12-15 W of electrical power
 - Side by side testing showed nearly comparable efficiency (vs. VCBB Rev1), nuclease inactivation, and biosafety; VCBB Rev3 now outperforming ultrasonic
- For processing 4 sample tubes, ultrasonic and VCBB use about the same total energy and time, but US is 1 at a time, vs. 4 at a time for VCBB
- **Primary reasons for using VCBB for NAATOS are performance and cost, as US BOM is still > \$500**
- Will continue to monitor price-lowering efforts and freely share our US lysis knowledge generated




Further look at key requirements / specs




Vertical (upright) tube orientation is required

- Recent designs with upright tube orientation have shown consistently superior results to horizontal tube configurations, including at low copy numbers
- Allows for smaller heater area/volume with liquid collected at bottom of tube

Improved lysis efficiency:
Shaking tube in vertical position
N=13 individual experiments





Averaged bio-reps normalized to control




All else constant, throw appears to have the largest impact on lysis efficiency

- With integrated heating, current best spec for throw is 7 +/- 0.5mm
- Will perform additional tolerance testing with current design in coming weeks




Hot start (but not integrated heating) experiment showed fairly comparable results for 6.5 to 8 mm throw

- Non-heated lysis experiments required throw >8mm for best efficiency
- We turned throw up to 11 (mm) one time, but results were predictable given the *Spinal Tap* reference
- This hot start experiment showed modest improvement from 6.5mm to 8mm
- Rev3 design has 7mm throw, which has worked consistently well for integrated heating versions




Oscillation frequency has a smaller effect

- Current best spec with integrated heating is 65 +/- TBD Hz
 - From a lysis efficiency perspective, tolerance is likely around +/- 5 Hz
 - From a production / efficiency perspective, tolerance more likely +/- 1 or 2 Hz
- Previous testing largely a result of designing different flexure plates with lower or higher resonant frequencies, ranging from 50 to 110 Hz
- Want to minimize power draw via lower frequencies, but still need some minimum value for a given throw and time regime, which have larger impacts
- No commercial devices for lysis use < 50 Hz





Heating temp tolerance is moderate

- Current best spec is 95 +/- 3C at the heater/tube interface, assuming a 3-5C differential between tube exterior and liquid
- Basically want to get all the liquid to 90C without boiling it
- Plot to right shows thermocouple traces immersed in a sample tube inserted in VCBB heater (red traces) vs. an aluminum heat block (black traces), both nominally at 95C




Heating temp driven more by nucleases than lysis

- Lysis efficiency does not degrade at temps lower than 95C – upper right fig.
 - Sample holder *should* (shall?) be insulative enough to maintain temp of at least 65C during shaking
- Nuclease inactivation suffers at temps less than 95C (lower right fig), so sticking with current program of holding 95C for 2 minutes
 - Have not explored lower temp for longer duration – energetically didn't see an advantage. Also easier to keep temp up during shaking by heating higher/shorter.
- Biosafety confirmed for current program (showed images before)




Durations and timing of heat

- Initial heating ramp is balance between power and time
 - Derived requirement from available time, battery considerations, etc.
 - Currently requires 1 to 2 minutes to reach 95C depending on starting and ambient conditions
- Duration of holding at 95C (2 minutes) driven by nuclease and biosafety considerations rather than lysis efficiency
 - Plot on right shows no to minimal lysis efficiency gains for longer heating times prior to initiating bead beating



Durations and timing of shaking (bead beating)

- As shown previously (slide 11), no lysis efficiency boost by keeping heater on while shaking vs. switching off heater
 - Assuming holder at least as insulating as current design
 - Slide 21 showed lysis efficiency equivalency for temps down to 65C
 - VCBB ends the 4 minute shake in the 60-70C range, so stays within optimal lysis zone
- Longer shaking provides improved efficiency with decreasing gains
- **In prototypes, want to make timings adjustable to optimize for downstream NAATOS assay modifications**



Power and battery considerations

- Key requirement is the battery needs to run for 24 tubes w/out a charge, and be rechargeable worst case overnight
- Current heating demands
 - Peak ~35 W during max 2 min ramp from a cold start
 - Steady state heating ~7W for 2 min hold
 - Shaking ~9W max for 4 mins
 - Total 2 Wh per run
 - If efficiently do 6 runs of 4 tubes each, total of 12 Wh per day (1-2 18650 cells)
 - Very unlikely worst-case scenario of 1 tube at a time → 48 Wh per day (5 18650 cells)
 - Probably ruled out by other battery requirements – will need to determine what a realistic worst-case scenario looks like
- Other battery considerations roll into related requirements on total cost, shipping cost and restrictions, etc.

Will select a single sample tube for compatibility

- Most work to date used a single 2mL tube that is not “squeezy” and does not have a dropper top
- NAATOS requires a squeezy tube with a secondary dropper top
- Determined polypropylene is the best material, now testing various options (examples on next slide)
- Will determine final tube in next ~month, but very unlikely to have diameters outside the 8-12 mm range or mass outside 1.5 to 3.4 g
- Use 150mg of 100um diameter glass beads in the tube
- Break off swab tip in tube and leave it there for duration of sample prep
- When pick final tube, need to confirm no need for dummy tubes for weight balancing purposes

Sample tube dimensions



1.5 g




3.4 g




2.3 g

Traditional 2 mL gasketed BB tube




~10 mm

“Wide” tube



~12 mm


“Narrow” tube




~8 mm

Dropper tube performance assessment

- Air coupled heater built to investigate alternative tube sizes in lab
 - Trades time to heat for ease of adaptation for new tube sizes
- Early results seem to indicate that minor variations in tube geometry do not change efficiency substantially



Alternate tube lysis efficiency



User interface considerations



- 1 step, walkaway operation for final unit; prototypes do not have to be
- For final unit, no user-accessible parts or maintenance
- For development – control via updating software or user interface (can be simple computer command line interface, doesn't have to be onboard)
- Indicator(s) of progress TBD
 - Battery/charging status
 - Light colors / blinking vs. simple digital display
 - Indicate heating, shaking, done, still hot vs. just in progress and done
- May need additional “lock out” period to let tube cool before user accesses it
 - User safety
 - Ensuring hot liquid doesn't damage NAATOS device/enzymes

Current design details





Voice Coil (History: what is it now)

- Path finding unit to explore 3-10mm displacement at 50-80Hz frequencies
- As noted above, initially found efficacy at ~8mm displacement ~70Hz
- Crash task to get lab hardware into the field and add simultaneous heating
- Uganda unit (Rev1) had time pressure to get into the study
 - Unit performed well in the study before encountering a failure (detailed on subsequent slide)



Voice Coil (History: what is it now)


- During Uganda testing, development continued
- Significant improvements to both the heater and voice coil
 - Three plate flexure and support
 - Improve heater block
 - General clean up
 - Switch to vertical tubes
- Found significant lysis efficiency improvement (>20%) with vertical orientation
- Used some of the improvement to lower frequency and displacement (margin improvement)
- Currently at 7mm @ 64Hz
- Many existing design elements are driven by flexibility prototype-ability and fast turn




Voice Coil (Flexure plate geometry optimization)



Off the shelf: insufficient



linear spiral: max 800 Mpa



Tuned spiral: max 400 Mpa

- Distribute the stress on the flexure arms
- Improve displacement to 8mm
- Decreased peak stress by ~3x using geometry alone

Voice Coil (material selection)

- How much stress relief do we need
 - For extremely long lifetimes we need to be under 600 MPa peak
- Going to stick with 1095 for the time being
 - Some SS are marginally better but with much less availability
 - Titanium is interestingly better, but I don't want to make that jump yet
 - Keeping material as reserve margin




Figure 4
AISI 1095. The effect of pits on the fatigue life. The numbers refer to the pit cross section area in μm^2 .


Voice Coil (material selection)

Material	Elastic Modulus (GPa)	Yield Strength (MPa)	Ultimate Strength (MPa)	Reversed Bending Fatigue Limit(MPa)
7C27Mo2	210	1300	1800	710
Cronidur 30	210	1850	2150	1141
XD15NW	198	1580	2350	1095
100Cr6	207	2033	2240	938
C300	189	1971	2020	861
C350	199	2316	2358	758
AerMet100	194	1689	1965	945
AerMet310	192	1900	2170	1030
AerMet340	192	2160	2430	986
M50	189	2200	2480	1010


- More conventional grades may be a good option (420, 316L, etc.) some evidence that they could be close to filling requirements, but good data was not readily available and was not a hole we needed to go down to achieve our experimental objectives

Voice Coil (need more displacement)

- Final stress optimized design
- Exploit nonlinearity of 2nd moment of area



- Thinner plates decrease stress in individual plates
- Combining multiples keeps resonance frequency up
- Multiple spaced plates improve the centering of the coil in the yoke
- Three 0.025in 1095 Spring steel plates




Voice Coil (COTS parts)

- Leveraged off-the-shelf surface resonator voice coil
- Samtronic, Huasheng (Shenzhen) Technology Limited
 - They will supply just the magnet and coil cup for \$22 US
 - Min order: 3k
 - They are interested in making the flexure as well (~2k tooling) this is only preliminary talk, and I am still baselining laser cutting




Voice Coil (build)






Voice Coil (build)

- Off-the-shelf coil
- Three Laser cut flexure plates
- Plastic spacers and screws
- Tested for 4 days of continuous running (22 million cycles)




Heater (original concept)

- Use FR4 for both the heater and the insulation
- Stack on identical PCBs
- Each plate forms a circuitous path for the escaping heat
- Plates stack up like laminations
- Current passes from one plate to the next like batteries





Heater (original concept)

- Using FR4 path provides good thermal insulation
- 45C delta




Heater (reduction to prototype)

- Simple stack of identical layers
- Forms both a heater and insulation and the body of the unit
- FR4 is cheap
- Currently using 0.8 mm double-sided boards for ease and flexibility
- Wanted to use 4 and 6-layer boards to decrease the number of layers but was becoming a headache





Control and drivers

- All on board
 - Microcontroller
 - Class D amplifier
 - PWM PID heater control
 - Thermistor feedback
 - Almost no wiring
- The next version could have no wiring
- Talk about ferrite vs inductor filter
- Heater insulation and control all weigh **80g**




Components




Reflow assembly technique

- Low temp solder paste melt temp 138C
- Stack the layers in a jig for alignment and clamping
- Drive current through the heater layers to melt the solder
- Let cool and you have an assembled heater
- Using the heater to assemble itself... priceless




Monolithic heater assembly

- Option one
 - Glue laminated structure
 - Use thermal cure epoxy during the coil clamping and soldering process
- Option 2
 - The PCB surfaces are tightly held together in the current design
 - Conformal coating could be applied to all laminated surfaces






Bringing it all together

- Current voice and heater situation




Bringing it all together

- Coupled oscillator suspension
- Vertical structure




Failure analysis of Uganda unit (Rev1)

- Observable rubbing of the coil on the yoke
- Unit worked properly after reassembly
- Explains observed behavior
- Flexure plates were intact



Failure analysis of Uganda unit (Rev1)

- Uganda-era units were a bit flexible transverse to the axis of vibration
- Allowed some $m=1, n=1$ drum mode
- This was improved before the failure was observed
- Multilayer flexure and better plate support significantly improve transverse rigidity



Ran out of time/future considerations

- Might be worth doing a custom design on the voice coil
- Boards need conformal coatings
- Time to move to stainless for flexure
- Reflow solder the heater plates together
- Currently using 0.8 mm two-layer boards move to 4 or 6 layers.
- Stack needs conformal coating and/or lamination.
- Route all connections from board to board
- Accelerometer on board to enable closed-loop operation
- ~2000/year is not a lot
- Design is still optimized for flexibility and rapid prototyping
 - Think the basic VC structure is solid
 - The heater may or may not be ideal for this target volume.
 - Maybe more COTS parts
 - Polyimide flex heaters
- Can have zero moving wires by having the UI on the device and supplying 12v on the innermost plate and ground on the outer.

Initial BOM analysis

Key parts	Min quantity	Price	Confidence	Note
VCA (cup and magnet)	3000	\$22	solid	
PCB stack for heater and rack	40	\$0.6	solid	
Copper shim	100	\$0.4	solid	
Tube clamp	1000	\$0.5	ROM	
Rack to VAC mount	1000	\$2	ROM	
Stand offs	100	\$0.2	solid	
Nuts and bolts	100	\$0.5	ROM	
Class D amplifier 18W (TPA3138)	1000	\$0.4	solid	
SamD21 (ATSAMD2118A)	1000	\$4	solid	In normal times: \$1 we can do better
Other electronics parts and PCB	1000	\$5	ROM	
Vibration isolation	1000	\$3	ROM	
All else – Case, SPU, battery	TBD	<\$50	ROM	
Total:		<\$100	ROM	

Plan moving forward

- Similar to power module, want to push toward cheapest possible solution first
- Schedule and milestones
- Meeting cadence

NAATOS sample prep module requirements



Selected Requirements

ID	System requirements
240	Sample prep module will be produced at 2000 units/year quantities.
165	Sample prep module will comply IEC 61010-1.
226	Sample prep module will comply with IEC 61010-2-101.
229	Sample prep module will comply with IEC 61326-1.
230	Sample prep module will comply with IEC 61326-2-6.
246	Sample prep module will comply with ISO 18113-1 for labeling that provides necessary information for proper use, handling, and identification.
169	Sample prep module will comply with transportation regulations related to electronic devices containing batteries.
242	Sample prep module will comprise assembled raw materials available from more than one global supplier.
205	Sample prep module will cost \$200 EXW at 1000 units/year.
228	Sample prep module will fit on a table within a space \leq 30 cm deep, 45 cm wide, and 45 cm high.
180	Sample prep module will have an MTTF of 12,000 cycles -runs.
210	Sample prep module will include a stand-alone power source.
235	Sample prep module will indicate or enforce correct orientation for loading sample prep tube.
254	Sample prep module will minimize the burn risk to a user.
214	Sample prep module will render a sample biosafe during operation.
231	Sample prep module will support \geq 24 tests during a power outage.
255	Sample prep module will visually communicate progress and completion of sample prep process.

Selected Requirements

ID	System requirements
206	Sample prep tube (with reagents) will cost \$0.15 EXW TBR at 10 million units/year.
212	Sample prep tube must contain stabilized lysis reagents for transport and storage.
186	Sample prep tube will comply with ISO 18113-1 for labeling that provides necessary information for proper use, handling, and identification.
245	Sample prep tube will comprise assembled raw materials available from more than one global supplier.
190	Sample prep tube will contain all reagents preloaded in stable form.
188	Sample prep tube will function as intended within 60 minutes after being unpackaged.
189	Sample prep tube will have a minimum shelf life of 18 years.
191	Sample prep tube will not require precise measurements to dispense correct lysate amount into test device.
219	Sample prep tube will provide printed space for operator to write pertinent clinical notes (e.g., patient ID).
237	Sample prep tube will remain closed during lysis in sample prep module.
192	Sample prep tube will require addition of ≤ 1 external reagent prior to insertion into Sample Prep module.
166	System prep module will require one interaction period to lyse sample.
173	System will maintain function at temperatures between 5–45°C.
171	System will meet ASTM D7386-16-“Standard Practice For Performance Testing Of Packages For Single Parcel Delivery Systems” transport standards.
172	System will produce a result within 60 minutes. Sample prep procedure will take < 15 minutes total including sample collection
176	System will provide printed IFU (pictorial with English instruction) and on-device labeling (FprEN ISO 18113-1, FprEN ISO 18113-2)
174	System will use ISO/IEC spec for symbols (FprEN ISO 18113-1, FprEN ISO 18113-2).
177	System will weigh < 1 kg. More like hand-portable and stable on benchtop while running or sliding the module on benchtop
227	Tongue swab will remain in sample prep tube after sample collection.

Additional not-yet-official requirements / specs

ID	Specification
TBD	The sample prep module will accommodate up to 4 sample tubes at a time
TBD	The sample prep module will be operable while charging the battery
TBD	The sample prep module battery will be fully rechargeable overnight
TBD	The sample prep module will produce lysis efficiency at least as high as the reference assay of heat block + BioSpec bead beater (TBR for NAATOS assay)
TBD	The sample prep module will inactivate nuclease activity ... TBD
TBD	The sample prep module will hold sample tubes in an upright orientation (angle tolerance TBD)
TBD	The sample prep module will provide a peak to peak displacement of the sample tube of 7 +/- 0.5 mm in the horizontal direction
TBD	The sample prep module will oscillate the sample tube at a frequency of 65 +/- TBD Hz
TBD	The sample prep module will heat the sample tube to a maximum of 95 +/- 3 degrees Celsius
TBD	The sample prep module should reach 95C within 2 minutes of a cold start
TBD	The sample prep module will hold the sample tube at 95C for 2 minutes before bead beating starts. This duration will be adjustable for prototype phase.
TBD	The sample prep module will not require simultaneous heating and bead beating to meet performance requirements
TBD	The sample prep module will perform bead beating for up to 4 minutes. This duration will be adjustable for prototype phase.
TBD	The sample prep module will maintain the sample prep tube's temperature greater than TBD C during the duration of bead beating
TBD	The sample prep module will be compatible with a given sample tube, specifically TBD with TBD characteristics
TBD	The sample prep module will enable walkaway operation
TBD	The sample prep module will not require any user maintenance or allow user modifications to operation (hardware or software)
TBD	During the prototype phase, the control software will be updatable via a standard computer connection or on-screen display

NAATOS overall requirements



Selected Requirements

ID	Product requirements
43	NAATOS TB V1 must operate within the humidity range of 70% to 90% non-condensing humidity without performance degradation.
44	NAATOS TB V1 must operate within a temperature range of +5 °C to +45 °C without performance degradation.
47	NAATOS TB V1 cost per test must not exceed \$5 USD.
62	NAATOS TB V1 must not rely on mains power during test operation.
65	NAATOS TB V1 must be designed to effectively protect against dust and contaminants, ensuring its functionality and integrity.
68	NAATOS TB V1 must visually signal that the device is performing as intended to communicate progress and completion of various steps and processes.
69	NAATOS TB V1 must be a hand transportable product that weighs less than < 1 kg.
71	NAATOS TB V1 must have a cycle life of 12,000 tests.
78	NAATOS TB V1 must have a shelf life of ≥24 months for reusable modules and ≥18 months from consumables.
80	NAATOS TB V1 must support the required daily throughput in the event of power outages.
85	NAATOS TB V1 must minimize the potential for a burn risk.
86	NAATOS TB V1 must be registered for in vitro diagnostic use through WHO Prequalification or other regulatory authority of equal or greater stringency.
95	NAATOS TB V1 must be able to run TBD tests in parallel to support a daily throughput of 10-25 tests per 6-hour day at L0, L1, or L2 health setting.
98	NAATOS TB V1 must operate without the need for maintenance during its intended lifespan.
99	NAATOS TB V1 must operate without the need for calibration during its intended lifespan.
63	NAATOS TB V1 must fit on a table within a space ≤ 60 cm deep, 90 cm wide, and 45 cm high.
50	NAATOS TB V1 must not require disposal of infectious waste.

Selected Requirements

ID	Product requirements
38	NAATOS TB V1 must require minimal user interaction periods (≤ 3) before walking away and returning to record the result (excluding sample prep).
46	NAATOS TB V1 must be able to perform more than one sample at the same time.
67	NAATOS TB V1 must include designated space to accurately record and display patient identifiers.
94	NAATOS TB V1 must tolerate $\pm 33\%$ variability of lysed sample volume introduced by user into test device.
61	NAATOS TB V1 must be ready to use without the need for additional components or accessories.
68	NAATOS TB V1 must visually communicate progress and completion of system processes.
81	NAATOS TB V1 must not require precise measuring of volume or time required for any step and should instead require number of drops or visual markers.
83	NAATOS TB V1 must support testing with a noninvasive tongue swab.
102	NAATOS TB V1 must use no more than two external reagents for sample preparation.
99	NAATOS TB V1 must operate without the need for calibration during its intended lifespan.
98	NAATOS TB V1 must operate without the need for maintenance during its intended lifespan.
66	NAATOS TB V1 must prevent the escape of amplifiable material into the testing area.
75	NAATOS TB V1 must achieve a clinical specificity $\geq 98\%$.
74	NAATOS TB V1 must achieve a clinical sensitivity $\geq 80\%$.
53	NAATOS TB V1 invalid rate must not exceed 2%.
63	NAATOS TB V1 must withstand resistance to shocks and vibrations from drops by users of 90 cm TBR height without performance failures.
248	NAATOS TB V1 must indicate to users that the test is invalid.
85	NAATOS TB V1 must minimize the burn risk to a user.

GH+
Labs

Thank you
—

